


Patient Quickcard Invia® Liberty™

Switch on



Press quickly

Switch off




Press for
3 seconds

Press only once

Pump needs
approx. 10 sec to
switch off

Alarm sounds

1




Problem description appears on display

2



Press both buttons simultaneously

3



Troubleshooting info appears on display

4

Solve problem yourself or consult your contact person